

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 006785  
12:49 27/12/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:45  
PRESSURIZING 2:47  
STERILIZING 3:57  
136.4 C 220kPa 3:57  
136.2 C 221kPa 4:27  
136.4 C 223kPa 4:57  
136.4 C 222kPa 5:27  
136.4 C 223kPa 5:57  
136.4 C 222kPa 6:27  
136.2 C 224kPa 6:57  
136.2 C 221kPa 7:27  
136.2 C 221kPa 7:27  
VENTING 7:27  
AIR DRYING 8:18  
CYCLE COMPLETE 38:18

-----  
Digital Signature #  
E87941E5D603EE98  
-----