
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001576
15:10 08/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:41
STERILIZING 3:48
136.4 C 225kPa 3:48
136.4 C 225kPa 4:18
136.4 C 224kPa 4:48
136.3 C 224kPa 5:18
136.4 C 225kPa 5:48
136.4 C 221kPa 6:18
136.4 C 224kPa 6:48
136.4 C 222kPa 7:18
136.4 C 222kPa 7:18
VENTING 7:18
AIR DRYING 8:09
CYCLE COMPLETE 8:19

Digital Signature #
263A7C658FA89119
