

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3818

Name: RITTER2  
Sterilizer ID: V2313228  
07/26/2024 01:27 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 127.2 F 0.0  
2:00 129.7 F 1.3  
4:00 137.1 F 3.8  
6:00 174.2 F 7.4  
8:00 191.6 F 11.5  
10:00 210.4 F 5.2  
12:00 221.4 F 9.5  
14:00 230.7 F 10.7  
16:00 235.0 F 10.6  
18:00 234.9 F 8.3  
20:00 246.6 F 14.4  
22:00 257.4 F 20.3  
24:00 267.5 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.4  
1:00 273.1 F 30.3  
2:00 272.8 F 30.2  
3:00 272.5 F 30.3  
4:00 272.0 F 30.3

Min 270.9 F 29.3  
Max 273.2 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.3 F 0.0  
5:00 203.3 F 0.0  
10:00 209.9 F 0.0  
15:00 208.7 F 0.0  
20:00 206.9 F 0.0  
25:00 205.4 F 0.0  
30:00 205.2 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 25:30  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:04:14  
CYCLE COMPLETE

07/26/2024 02:31 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_