
Midmark M11 - v1.0.0
Total Cycles: 5389

Name: RITTER 1
Sterilizer ID: V2178198
03/05/2025 01:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.1 F 0.0
2:00 169.0 F 3.7
4:00 179.5 F 7.8
6:00 201.1 F 12.6
8:00 218.5 F 6.5
10:00 231.3 F 11.4
12:00 231.4 F 7.9
14:00 244.0 F 13.5
16:00 242.1 F 11.9
18:00 254.0 F 17.9
20:00 265.1 F 24.2
22:00 271.0 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.4 F 29.4
2:00 272.1 F 29.7
3:00 272.2 F 29.3
4:00 272.2 F 29.8

Min 271.6 F 29.1
Max 272.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.7 F 0.0
5:00 203.7 F 0.0
10:00 206.9 F 0.0
15:00 204.1 F 0.0
20:00 205.6 F 0.0
25:00 214.0 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 22:41
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:01:46
CYCLE COMPLETE

03/05/2025 02:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____