
Midmark M11 - v1.0.0
Total Cycles: 5284

Name: RITTER 1
Sterilizer ID: V2178198
01/28/2025 10:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.7 F 0.0
2:00 72.3 F 0.0
4:00 94.2 F 0.1
6:00 132.2 F 2.2
8:00 152.8 F 5.0
10:00 172.1 F 8.5
12:00 191.5 F 12.0
14:00 202.7 F 3.0
16:00 215.7 F 6.6
18:00 228.5 F 10.3
20:00 233.8 F 8.7
22:00 237.0 F 10.3
24:00 237.8 F 8.5
26:00 246.6 F 13.7
28:00 257.7 F 19.7
30:00 268.0 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.8 F 29.2
2:00 271.8 F 28.4
3:00 272.8 F 29.3
4:00 272.4 F 28.7

Min 271.6 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 192.1 F 0.0
10:00 198.5 F 0.0
15:00 204.3 F 0.0
20:00 207.9 F 0.0
25:00 209.1 F 0.0
30:00 211.7 F 0.0
Drying Complete

FILLING: 0:41
HEATING: 31:22
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:09:39
CYCLE COMPLETE

01/28/2025 11:14 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____