
Midmark M11 - v1.0.0
Total Cycles: 4478

Name: RITTER 1
Sterilizer ID: V2178198
04/22/2024 03:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 91.2 F 0.0
2:00 92.8 F 0.7
4:00 111.6 F 2.5
6:00 140.2 F 4.9
8:00 213.5 F 8.1
10:00 190.8 F 11.9
12:00 215.3 F 3.8
14:00 221.7 F 7.4
16:00 231.0 F 11.3
18:00 231.3 F 6.5
20:00 241.2 F 12.0
22:00 238.6 F 9.6
24:00 250.6 F 15.5
26:00 260.7 F 21.4
28:00 271.0 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 271.4 F 29.0
2:00 271.6 F 29.1
3:00 271.6 F 29.2
4:00 271.4 F 29.1

Min 271.2 F 28.7
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.8 F 0.0
5:00 238.3 F 0.0
10:00 222.4 F 0.0
15:00 208.3 F 0.0
20:00 202.8 F 0.0
25:00 214.6 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 29:21
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:07:59
CYCLE COMPLETE

04/22/2024 04:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____