
Midmark M11 - v1.0.0
Total Cycles: 2317

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2022 09:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 215.6 F 0.0
2:00 216.8 F 3.7
4:00 221.6 F 9.0
6:00 227.0 F 6.8
8:00 234.7 F 10.6
10:00 236.1 F 7.2
12:00 243.1 F 11.7
14:00 244.4 F 12.7
16:00 257.2 F 19.5
18:00 268.5 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.6 F 29.6
2:00 272.0 F 28.9
3:00 271.3 F 28.8
4:00 272.6 F 29.6

Min 271.1 F 28.5
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.0 F 0.0
5:00 228.1 F 0.0
10:00 238.9 F 0.0
15:00 232.8 F 0.0
20:00 218.2 F 0.0
25:00 214.0 F 0.0
30:00 213.9 F 0.0
Drying Complete

FILLING: 2:45
HEATING: 19:12
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:08
CYCLE COMPLETE

09/13/2022 10:43 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____