
Midmark M11 - v1.0.5
Total Cycles: 2736

Name: RITTER2
Sterilizer ID: V2313228
08/04/2023 04:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 108.6 F 0.0
2:00 109.2 F 0.0
4:00 123.3 F 0.0
6:00 169.3 F 0.3
8:00 189.5 F 3.6
10:00 207.0 F 6.8
12:00 216.9 F 10.5
14:00 224.4 F 7.9
16:00 231.9 F 9.7
18:00 237.5 F 10.5
20:00 242.2 F 12.3
22:00 241.3 F 11.3
24:00 253.3 F 17.3
26:00 263.9 F 23.7
28:00 271.9 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.5 F 29.3
2:00 271.5 F 29.0
3:00 272.4 F 29.3
4:00 271.4 F 28.7

Min 271.4 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.9 F 0.0
5:00 216.1 F 0.0
10:00 217.9 F 0.0
15:00 212.7 F 0.0
20:00 209.4 F 0.0
25:00 208.5 F 0.0
30:00 209.9 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 28:08
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:06:41
CYCLE COMPLETE

08/04/2023 05:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____