
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001463
10:48 06/04/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:36
PRESSURIZING 3:36
STERILIZING 4:35
136.5 C 222kPa 4:35
136.4 C 225kPa 5:05
136.4 C 224kPa 5:35
136.4 C 225kPa 6:05
136.1 C 220kPa 6:35
136.1 C 223kPa 7:05
136.4 C 225kPa 7:35
136.3 C 222kPa 8:05
136.3 C 222kPa 8:05
VENTING 8:05
AIR DRYING 9:01
CYCLE COMPLETE 9:02

Digital Signature #
0D03E810F2731892
