
Midmark M11 - v1.0.0
Total Cycles: 3272

Name: RITTER 1
Sterilizer ID: V2178198
05/24/2023 07:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.6 F 0.0
2:00 78.7 F 0.0
4:00 85.1 F 0.0
6:00 110.1 F 0.0
8:00 140.6 F 0.7
10:00 169.2 F 4.0
12:00 187.2 F 7.4
14:00 204.2 F 11.1
16:00 220.2 F 5.3
18:00 227.3 F 8.3
20:00 239.3 F 12.6
22:00 236.7 F 9.5
24:00 239.0 F 9.4
26:00 247.2 F 13.7
28:00 258.7 F 20.0
30:00 269.3 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.5 F 29.0
2:00 271.6 F 28.1
3:00 272.2 F 28.9
4:00 272.2 F 28.6

Min 271.3 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.6 F 0.0
5:00 220.9 F 0.0
10:00 235.3 F 0.0
15:00 228.6 F 0.0
20:00 225.6 F 0.0
25:00 225.7 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:06
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:09:06
CYCLE COMPLETE

05/24/2023 08:42 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____