

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3190

Name: RITTER 1  
Sterilizer ID: V2178198  
05/02/2023 03:06 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 208.7 F 0.0  
2:00 197.4 F 3.3  
4:00 213.5 F 8.7  
6:00 232.0 F 9.4  
8:00 237.6 F 10.9  
10:00 235.6 F 8.9  
12:00 239.3 F 9.5  
14:00 249.2 F 15.0  
16:00 261.8 F 22.1  
18:00 272.8 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.5  
1:00 272.4 F 29.2  
2:00 272.2 F 28.8  
3:00 271.7 F 28.8  
4:00 272.7 F 29.2  
  
Min 271.5 F 28.4  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.2 F 0.0  
5:00 234.6 F 0.0  
10:00 230.6 F 0.0  
15:00 227.7 F 0.0  
20:00 228.7 F 0.0  
25:00 227.4 F 0.0  
30:00 229.5 F 0.0  
  
Drying Complete

FILLING: 2:43  
HEATING: 18:13  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:13  
CYCLE COMPLETE

05/02/2023 04:04 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_