
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001091
13:02 25/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:23
PRESSURIZING 3:27
STERILIZING 4:43
136.0 C 221kPa 4:43
136.3 C 223kPa 5:13
136.3 C 223kPa 5:43
136.3 C 223kPa 6:13
136.1 C 221kPa 6:43
136.4 C 224kPa 7:13
136.1 C 223kPa 7:43
136.1 C 223kPa 8:13
136.1 C 223kPa 8:13
VENTING 8:13
AIR DRYING 9:04
CYCLE COMPLETE 18:05

Digital Signature #
200271E1869FE28C
