
Midmark M11 - v1.0.0
Total Cycles: 4524

Name: RITTER 1
Sterilizer ID: V2178198
05/03/2024 02:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.0 F 0.0
2:00 111.2 F 0.4
4:00 126.7 F 2.6
6:00 156.1 F 5.5
8:00 182.1 F 8.4
10:00 201.4 F 12.5
12:00 212.9 F 4.7
14:00 220.5 F 8.8
16:00 230.9 F 13.0
18:00 233.8 F 9.0
20:00 240.8 F 11.6
22:00 243.1 F 12.0
24:00 254.0 F 17.6
26:00 264.2 F 23.6
28:00 272.1 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.5 F 29.2
2:00 271.9 F 28.6
3:00 271.6 F 28.8
4:00 272.4 F 29.2

Min 271.3 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.0 F 0.0
5:00 210.3 F 0.0
10:00 207.2 F 0.0
15:00 213.9 F 0.0
20:00 226.6 F 0.0
25:00 230.8 F 0.0
30:00 231.7 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 28:12
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:07:03
CYCLE COMPLETE

05/03/2024 03:43 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____