
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001205
15:21 03/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:51
PRESSURIZING 3:52
STERILIZING 5:05
136.4 C 224kPa 5:05
136.2 C 223kPa 5:35
136.3 C 223kPa 6:05
136.4 C 225kPa 6:35
136.4 C 225kPa 7:05
136.4 C 225kPa 7:35
136.4 C 225kPa 8:05
136.3 C 223kPa 8:35
136.3 C 223kPa 8:35
VENTING 8:35
AIR DRYING 9:29
CYCLE COMPLETE 20:18

Digital Signature #
FCC62A6BF66B24A7
