
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006311
8:18 18/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:04
STERILIZING 3:13
136.4 C 220kPa 3:13
136.3 C 222kPa 3:43
136.4 C 223kPa 4:13
136.4 C 224kPa 4:43
136.4 C 220kPa 5:13
136.4 C 224kPa 5:43
136.1 C 222kPa 6:13
136.2 C 222kPa 6:43
136.2 C 222kPa 6:43
VENTING 6:44
AIR DRYING 7:34
CYCLE COMPLETE 9:20

Digital Signature #
E242297AFE2B329D
