
Midmark M11 - v1.0.0
Total Cycles: 1480

Name: RITTER 1
Sterilizer ID: V2178198
12/21/2021 09:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.6 F 0.0
2:00 175.2 F 2.0
4:00 193.3 F 5.4
6:00 210.0 F 9.6
8:00 214.3 F 6.5
10:00 227.9 F 8.7
12:00 239.2 F 12.1
14:00 238.1 F 10.7
16:00 235.6 F 8.6
18:00 249.5 F 15.3
20:00 261.2 F 21.9
22:00 271.9 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.6 F 29.4
2:00 272.3 F 28.9
3:00 271.5 F 28.7
4:00 272.6 F 29.5

Min 271.4 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.6 F 0.0
5:00 214.5 F 0.0
10:00 226.9 F 0.0
15:00 230.9 F 0.0
20:00 228.3 F 0.0
25:00 225.5 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 22:36
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:01:02
CYCLE COMPLETE

12/21/2021 10:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____