
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001463
10:51 20/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:07
PRESSURIZING 2:11
STERILIZING 3:14
136.4 C 223kPa 3:14
136.1 C 223kPa 3:44
136.4 C 226kPa 4:14
136.4 C 224kPa 4:44
136.2 C 223kPa 5:14
136.1 C 221kPa 5:44
136.2 C 223kPa 6:14
136.2 C 222kPa 6:44
136.2 C 222kPa 6:44
VENTING 6:44
AIR DRYING 7:34
CYCLE COMPLETE 7:40

Digital Signature #
BE6AA8B1C0CA834F
