
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002436
13:37 22/01/2025
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 4:44
STERILIZING 5:33
136.0 C 220kPa 5:33
136.2 C 223kPa 6:03
136.4 C 226kPa 6:33
136.4 C 223kPa 7:03
136.4 C 224kPa 7:33
136.3 C 223kPa 8:03
136.3 C 225kPa 8:33
136.4 C 225kPa 9:03
136.3 C 225kPa 9:33
136.4 C 222kPa 10:03
136.1 C 223kPa 10:33
136.1 C 222kPa 11:03
136.3 C 224kPa 11:33
136.4 C 226kPa 12:03
136.4 C 226kPa 12:33
136.4 C 226kPa 13:03
136.3 C 225kPa 13:33
136.4 C 224kPa 14:03
136.4 C 225kPa 14:33
136.4 C 223kPa 15:03
136.4 C 226kPa 15:33
136.4 C 226kPa 15:33
VENTING 15:33
AIR DRYING 16:30
CYCLE COMPLETE 17:02

Digital Signature #
284189DD4D3FAFC8
