
Midmark M11 - v1.0.5
Total Cycles: 674

Name: RITTER2
Sterilizer ID: V2313228
01/10/2022 09:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.0 F 0.0
2:00 76.6 F 0.0
4:00 102.8 F 0.0
6:00 119.5 F 0.4
8:00 142.1 F 3.1
10:00 166.1 F 6.1
12:00 185.0 F 9.5
14:00 202.9 F 13.4
16:00 215.2 F 6.7
18:00 228.1 F 11.0
20:00 229.3 F 7.1
22:00 239.8 F 12.5
24:00 239.1 F 10.7
26:00 251.0 F 16.7
28:00 261.4 F 22.8
30:00 271.2 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.5
1:00 272.6 F 29.8
2:00 272.1 F 30.2
3:00 271.8 F 29.4
4:00 272.5 F 30.1

Min 271.5 F 29.4
Max 272.6 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.0 F 0.0
5:00 207.4 F 0.0
10:00 221.3 F 0.0
15:00 218.1 F 0.0
20:00 214.1 F 0.0
25:00 225.0 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 30:38
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:08:53
CYCLE COMPLETE

01/10/2022 10:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____