

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1094

Name: RITTER2  
Sterilizer ID: V2313228  
05/11/2022 11:11 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.7 F 0.0  
2:00 185.7 F 4.0  
4:00 196.9 F 8.9  
6:00 221.7 F 8.4  
8:00 229.7 F 10.2  
10:00 232.8 F 8.0  
12:00 242.3 F 11.5  
14:00 245.8 F 13.4  
16:00 257.9 F 20.0  
18:00 269.4 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.4 F 29.4  
1:00 272.9 F 29.9  
2:00 272.1 F 29.7  
3:00 272.4 F 29.6  
4:00 272.6 F 30.0  
  
Min 271.9 F 29.3  
Max 273.1 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.3 F 0.0  
5:00 211.5 F 0.0  
10:00 230.2 F 0.0  
15:00 225.7 F 0.0  
20:00 222.3 F 0.0  
25:00 222.9 F 0.0  
30:00 222.1 F 0.0  
Drying Complete

FILLING: 2:05  
HEATING: 19:39  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:52  
CYCLE COMPLETE

05/11/2022 12:09 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_