
Midmark M11 - v1.0.5
Total Cycles: 2073

Name: RITTER2
Sterilizer ID: V2313228
02/22/2023 08:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.0 F 0.0
2:00 81.1 F 0.0
4:00 110.0 F 0.0
6:00 141.1 F 0.0
8:00 181.3 F 1.2
10:00 196.2 F 3.5
12:00 207.3 F 6.3
14:00 218.6 F 9.8
16:00 225.4 F 10.4
18:00 229.2 F 8.7
20:00 241.3 F 13.3
22:00 239.2 F 10.8
24:00 236.9 F 8.9
26:00 250.0 F 15.4
28:00 261.2 F 21.9
30:00 271.8 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.6 F 29.3
2:00 271.5 F 29.0
3:00 272.5 F 29.2
4:00 271.3 F 28.9

Min 271.3 F 28.6
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 201.1 F 0.0
10:00 206.2 F 0.0
15:00 219.2 F 0.0
20:00 221.2 F 0.0
25:00 229.1 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:33
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:08:33
CYCLE COMPLETE

02/22/2023 09:18 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____