
Midmark M11 - v1.0.5
Total Cycles: 3991

Name: RITTER2
Sterilizer ID: V2313228
10/02/2024 04:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.6 F 0.0
2:00 159.3 F 3.0
4:00 181.2 F 7.0
6:00 205.8 F 11.4
8:00 224.9 F 6.4
10:00 235.9 F 11.4
12:00 233.0 F 8.5
14:00 240.2 F 11.1
16:00 244.7 F 13.0
18:00 256.5 F 19.3
20:00 267.3 F 25.9
22:00 273.2 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.5
1:00 272.7 F 29.8
2:00 272.3 F 30.1
3:00 272.0 F 29.5
4:00 272.7 F 30.2

Min 271.3 F 29.4
Max 273.1 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.9 F 0.0
5:00 208.7 F 0.0
10:00 189.6 F 0.0
15:00 187.0 F 0.0
20:00 183.7 F 0.0
25:00 181.3 F 0.0
30:00 179.4 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 22:15
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:01:26
CYCLE COMPLETE

10/02/2024 05:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____