
Midmark M11 - v1.0.5
Total Cycles: 3521

Name: RITTER2
Sterilizer ID: V2313228
04/12/2024 10:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.7 F 0.0
2:00 164.5 F 1.9
4:00 186.4 F 5.9
6:00 206.2 F 10.4
8:00 223.9 F 6.6
10:00 233.1 F 10.1
12:00 233.7 F 8.6
14:00 243.2 F 13.4
16:00 242.8 F 12.4
18:00 255.2 F 18.8
20:00 266.3 F 25.6
22:00 272.7 F 30.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.7
1:00 272.9 F 30.3
2:00 271.3 F 29.6
3:00 272.1 F 30.5
4:00 272.1 F 30.0

Min 271.1 F 29.5
Max 273.1 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.1 F 0.0
5:00 224.1 F 0.0
10:00 208.0 F 0.0
15:00 184.4 F 0.0
20:00 179.9 F 0.0
25:00 177.1 F 0.0
30:00 176.0 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 22:27
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:01:30
CYCLE COMPLETE

04/12/2024 11:12 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____