
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000214
9:21 22/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:15
PRESSURIZING 3:21
STERILIZING 4:58
136.0 C 228kPa 4:58
136.2 C 232kPa 5:28
136.2 C 232kPa 5:58
136.2 C 231kPa 6:28
136.4 C 233kPa 6:58
136.1 C 231kPa 7:28
136.4 C 233kPa 7:58
136.1 C 231kPa 8:28
136.1 C 231kPa 8:28
VENTING 8:28
AIR DRYING 9:31
CYCLE COMPLETE 9:41

Digital Signature #
94BAB7D9A77BDBF9
