
Midmark M11 - v1.0.0
Total Cycles: 2231

Name: RITTER 1
Sterilizer ID: V2178198
08/16/2022 02:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.1 F 0.0
2:00 184.3 F 4.1
4:00 198.2 F 9.1
6:00 225.8 F 7.2
8:00 231.2 F 10.2
10:00 235.4 F 7.7
12:00 243.9 F 13.0
14:00 242.0 F 11.4
16:00 255.3 F 18.2
18:00 266.8 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 271.4 F 28.7
2:00 272.5 F 29.0
3:00 272.6 F 29.4
4:00 271.8 F 28.7

Min 270.9 F 28.1
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 217.4 F 0.0
10:00 230.5 F 0.0
15:00 230.0 F 0.0
20:00 227.2 F 0.0
25:00 230.8 F 0.0
30:00 229.0 F 0.0
Drying Complete

FILLING: 2:14
HEATING: 19:24
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:52
CYCLE COMPLETE

08/16/2022 03:01 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____