
Midmark M11 - v1.0.5
Total Cycles: 2081

Name: RITTER2
Sterilizer ID: V2313228
02/22/2023 05:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.5 F 0.0
2:00 199.0 F 3.5
4:00 217.3 F 9.1
6:00 228.3 F 7.7
8:00 237.8 F 11.8
10:00 238.2 F 10.3
12:00 238.2 F 9.7
14:00 253.6 F 17.4
16:00 266.4 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.3 F 28.9
2:00 272.7 F 29.4
3:00 271.7 F 28.5
4:00 272.7 F 29.3

Min 271.6 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.0 F 0.0
5:00 222.3 F 0.0
10:00 232.6 F 0.0
15:00 236.3 F 0.0
20:00 227.6 F 0.0
25:00 219.2 F 0.0
30:00 214.8 F 0.0
Drying Complete

FILLING: 2:28
HEATING: 17:26
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:02
CYCLE COMPLETE

02/22/2023 06:29 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____