
Midmark M11 - v1.0.5
Total Cycles: 3725

Name: RITTER2
Sterilizer ID: V2313228
06/21/2024 08:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 70.7 F 0.0
4:00 74.8 F 0.2
6:00 100.4 F 1.8
8:00 130.8 F 4.3
10:00 163.1 F 7.2
12:00 190.4 F 10.7
14:00 204.9 F 4.2
16:00 223.4 F 7.3
18:00 231.5 F 11.2
20:00 226.8 F 6.6
22:00 239.7 F 11.9
24:00 237.2 F 9.5
26:00 249.6 F 15.5
28:00 260.2 F 21.5
30:00 270.2 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.5 F 29.7
2:00 271.5 F 28.9
3:00 271.9 F 29.6
4:00 272.5 F 29.4

Min 271.0 F 28.7
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.3 F 0.0
5:00 224.4 F 0.0
10:00 179.3 F 0.0
15:00 175.0 F 0.0
20:00 169.9 F 0.0
25:00 168.0 F 0.0
30:00 167.5 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:58
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:09:03
CYCLE COMPLETE

06/21/2024 09:24 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____