
Midmark M11 - v1.0.5
Total Cycles: 279

Name: RITTER2
Sterilizer ID: V2313228
08/10/2021 02:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.7 F 0.0
2:00 167.6 F 3.0
4:00 183.2 F 7.6
6:00 206.7 F 13.2
8:00 226.0 F 8.4
10:00 235.6 F 11.7
12:00 240.8 F 12.2
14:00 241.8 F 11.5
16:00 255.1 F 18.4
18:00 266.6 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.9 F 30.4
2:00 271.7 F 29.5
3:00 273.0 F 30.3
4:00 271.8 F 29.5

Min 270.9 F 28.9
Max 273.2 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.9 F 0.0
5:00 222.4 F 0.0
10:00 229.4 F 0.0
15:00 227.8 F 0.0
20:00 217.3 F 0.0
25:00 207.2 F 0.0
30:00 197.5 F 0.0

Drying Complete

FILLING: 1:56
HEATING: 19:31
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:33
CYCLE COMPLETE

08/10/2021 03:42 PM
ACCEPTED: X
REJECTED:
Operator: U01