
Midmark M11 - v1.0.0
Total Cycles: 5254

Name: RITTER 1
Sterilizer ID: V2178198
01/17/2025 12:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.4 F 0.0
2:00 187.9 F 2.2
4:00 190.6 F 6.3
6:00 207.7 F 11.0
8:00 216.4 F 4.3
10:00 227.9 F 9.1
12:00 236.9 F 10.8
14:00 239.9 F 11.3
16:00 238.1 F 9.6
18:00 251.1 F 15.8
20:00 262.0 F 22.0
22:00 272.0 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.5 F 29.0
2:00 271.8 F 28.4
3:00 271.3 F 28.4
4:00 272.5 F 29.0

Min 271.2 F 28.1
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.4 F 0.0
5:00 203.1 F 0.0
10:00 231.0 F 0.0
15:00 225.1 F 0.0
20:00 216.0 F 0.0
25:00 213.7 F 0.0
30:00 212.9 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 22:34
STERILIZING: 4:00
VENTING: 3:53
DRYING: 30:00
TOTAL CYCLE: 01:01:36
CYCLE COMPLETE

01/17/2025 01:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____