
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002163
8:42 25/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 2:02
STERILIZING 3:07
136.5 C 220kPa 3:07
136.4 C 225kPa 3:37
136.4 C 224kPa 4:07
136.3 C 225kPa 4:37
136.4 C 226kPa 5:07
136.4 C 225kPa 5:37
136.4 C 225kPa 6:07
136.4 C 226kPa 6:37
136.4 C 226kPa 6:37
VENTING 6:37
AIR DRYING 7:31
CYCLE COMPLETE 25:34

Digital Signature #
A0311684159266EE
