
Midmark M11 - v1.0.5
Total Cycles: 4215

Name: RITTER2
Sterilizer ID: V2313228
12/13/2024 06:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.2 F 0.0
2:00 164.3 F 2.9
4:00 191.6 F 8.2
6:00 210.7 F 9.6
8:00 225.9 F 8.7
10:00 235.0 F 10.2
12:00 244.7 F 12.5
14:00 243.3 F 11.8
16:00 256.6 F 19.0
18:00 268.7 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.5 F 28.7
2:00 272.4 F 29.2
3:00 271.9 F 28.4
4:00 272.8 F 29.2

Min 271.5 F 28.2
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.0 F 0.0
5:00 239.7 F 0.0
10:00 224.3 F 0.0
15:00 214.9 F 0.0
20:00 205.0 F 0.0
25:00 197.0 F 0.0
30:00 190.9 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 19:03
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 00:58:31
CYCLE COMPLETE

12/13/2024 07:07 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____