
Midmark M11 - v1.0.5
Total Cycles: 193

Name: RITTER2
Sterilizer ID: V2313228
07/12/2021 10:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.7 F 0.0
2:00 186.1 F 3.7
4:00 198.4 F 9.1
6:00 219.8 F 9.1
8:00 227.3 F 11.5
10:00 236.2 F 9.8
12:00 237.3 F 9.3
14:00 249.9 F 15.5
16:00 261.5 F 22.1
18:00 272.0 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.8
1:00 271.9 F 29.7
2:00 272.0 F 29.7
3:00 271.6 F 29.8
4:00 271.7 F 30.1

Min 271.5 F 29.7
Max 272.7 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.3 F 0.0
5:00 235.8 F 0.0
10:00 227.2 F 0.0
15:00 230.5 F 0.0
20:00 226.0 F 0.0
25:00 230.4 F 0.0
30:00 224.8 F 0.0

Drying Complete

FILLING: 2:11
HEATING: 19:13
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:39
CYCLE COMPLETE

07/12/2021 11:41 AM
ACCEPTED: X
REJECTED:
Operator: U01