
Midmark M11 - v1.0.0
Total Cycles: 1092

Name: RITTER 1
Sterilizer ID: V2178198
08/18/2021 10:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.5 F 0.0
2:00 137.2 F 1.2
4:00 153.8 F 4.1
6:00 182.2 F 8.7
8:00 205.0 F 8.1
10:00 216.6 F 9.0
12:00 228.1 F 9.5
14:00 235.2 F 11.9
16:00 238.8 F 9.9
18:00 252.0 F 16.9
20:00 263.9 F 24.0
22:00 271.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.3 F 29.9
2:00 272.9 F 29.9
3:00 272.1 F 29.5
4:00 271.4 F 29.1

Min 271.1 F 29.0
Max 273.6 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.9 F 0.0
5:00 225.1 F 0.0
10:00 228.6 F 0.0
15:00 228.9 F 0.0
20:00 231.5 F 0.0
25:00 229.1 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 22:07
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 01:00:12
CYCLE COMPLETE

08/18/2021 11:27 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____