

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 006368  
9:20 27/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:20  
PRESSURIZING 2:24  
STERILIZING 3:27  
136.4 C 220kPa 3:27  
136.3 C 223kPa 3:57  
136.4 C 223kPa 4:27  
136.4 C 224kPa 4:57  
136.4 C 224kPa 5:27  
136.4 C 220kPa 5:57  
136.3 C 223kPa 6:27  
136.4 C 223kPa 6:57  
136.4 C 223kPa 6:57  
VENTING 6:57  
AIR DRYING 7:46  
CYCLE COMPLETE 37:46

-----  
Digital Signature #  
354E369CE98C0FDF  
-----