

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1263

Name: RITTER2  
Sterilizer ID: V2313228  
07/06/2022 02:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.8 F 0.0  
2:00 159.5 F 2.6  
4:00 171.9 F 6.1  
6:00 194.7 F 10.8  
8:00 212.4 F 4.6  
10:00 224.0 F 9.4  
12:00 232.1 F 10.0  
14:00 238.0 F 11.1  
16:00 235.8 F 9.0  
18:00 249.9 F 15.7  
20:00 261.3 F 22.2  
22:00 272.0 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.0  
1:00 272.3 F 29.9  
2:00 272.5 F 29.6  
3:00 271.6 F 29.0  
4:00 271.2 F 29.1

Min 271.1 F 28.8  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.9 F 0.0  
5:00 238.9 F 0.0  
10:00 190.8 F 0.0  
15:00 187.7 F 0.0  
20:00 186.3 F 0.0  
25:00 197.0 F 0.0  
30:00 194.3 F 0.0  
Drying Complete

FILLING: 1:09  
HEATING: 22:39  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:01:01  
CYCLE COMPLETE

07/06/2022 03:13 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_