
Midmark M11 - v1.0.5
Total Cycles: 1571

Name: RITTER2
Sterilizer ID: V2313228
10/06/2022 06:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.5 F 0.0
2:00 78.0 F 0.0
4:00 88.1 F 0.0
6:00 116.0 F 0.0
8:00 148.1 F 0.0
10:00 189.4 F 0.1
12:00 209.9 F 1.3
14:00 219.6 F 4.6
16:00 228.6 F 7.9
18:00 239.9 F 11.8
20:00 234.0 F 8.5
22:00 243.0 F 12.0
24:00 244.6 F 12.7
26:00 243.2 F 12.0
28:00 255.9 F 18.6
30:00 267.0 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.5 F 29.0
2:00 271.6 F 28.7
3:00 272.3 F 29.0
4:00 271.4 F 28.5

Min 271.4 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.5 F 0.0
5:00 214.0 F 0.0
10:00 195.9 F 0.0
15:00 201.9 F 0.0
20:00 182.8 F 0.0
25:00 187.2 F 0.0
30:00 179.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 31:28
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:09:36
CYCLE COMPLETE

10/06/2022 08:00 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____