
Midmark M11 - v1.0.0
Total Cycles: 3735

Name: RITTER 1
Sterilizer ID: V2178198
09/07/2023 01:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.1 F 0.0
2:00 179.3 F 2.1
4:00 192.1 F 6.1
6:00 208.6 F 11.1
8:00 222.2 F 5.5
10:00 233.2 F 10.5
12:00 233.0 F 7.9
14:00 244.8 F 13.0
16:00 244.2 F 12.5
18:00 256.4 F 18.9
20:00 266.9 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.4 F 29.2
2:00 271.4 F 28.4
3:00 272.3 F 29.2
4:00 271.9 F 28.7

Min 271.0 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.6 F 0.0
5:00 223.9 F 0.0
10:00 238.8 F 0.0
15:00 232.9 F 0.0
20:00 225.3 F 0.0
25:00 226.6 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 21:27
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:00:42
CYCLE COMPLETE

09/07/2023 02:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____