

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000774  
15:09 28/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:04  
PRESSURIZING 3:08  
STERILIZING 4:13  
136.4 C 232kPa 4:13  
136.3 C 232kPa 4:43  
136.3 C 232kPa 5:13  
136.3 C 232kPa 5:43  
136.4 C 230kPa 6:13  
136.3 C 232kPa 6:43  
136.4 C 233kPa 7:13  
136.4 C 233kPa 7:43  
136.4 C 233kPa 7:43  
VENTING 7:43  
AIR DRYING 8:41  
CYCLE COMPLETE 8:44

-----  
Digital Signature #  
B56BF99ED95B4241  
-----