
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
1.2uS / 0.7ppm
CYCLE NUMBER 003714
14:37 27/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:30
STERILIZING 3:25
136.5 C 225kPa 3:25
136.4 C 221kPa 3:55
136.3 C 225kPa 4:25
136.0 C 219kPa 4:55
136.4 C 223kPa 5:25
136.3 C 225kPa 5:55
136.4 C 222kPa 6:25
136.4 C 223kPa 6:55
136.4 C 223kPa 6:55
VENTING 6:55
AIR DRYING 7:47
CYCLE COMPLETE 7:51

Digital Signature #
3CDE64A29973F6D0
