
Midmark M11 - v1.0.5
Total Cycles: 2094

Name: RITTER2
Sterilizer ID: V2313228
02/27/2023 01:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.8 F 0.0
2:00 189.1 F 4.1
4:00 204.0 F 9.7
6:00 221.0 F 6.7
8:00 233.6 F 11.6
10:00 235.6 F 9.8
12:00 236.2 F 8.8
14:00 251.3 F 16.4
16:00 263.8 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.3 F 29.3
2:00 272.0 F 29.6
3:00 272.1 F 29.1
4:00 271.9 F 29.5

Min 271.4 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.2 F 0.0
5:00 223.9 F 0.0
10:00 238.4 F 0.0
15:00 228.9 F 0.0
20:00 224.5 F 0.0
25:00 222.9 F 0.0
30:00 222.7 F 0.0

Drying Complete

FILLING: 2:04
HEATING: 17:54
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:14
CYCLE COMPLETE

02/27/2023 02:51 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____