
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000007
12:12 13/12/2019
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:51
PRESSURIZING 1:03
STERILIZING 2:02
136.4 C 229kPa 2:02
135.9 C 224kPa 2:32
136.4 C 226kPa 3:02
136.0 C 221kPa 3:32
136.4 C 220kPa 4:02
136.4 C 225kPa 4:32
136.4 C 225kPa 5:02
136.2 C 222kPa 5:32
136.2 C 222kPa 5:32
VENTING 5:32
AIR DRYING 6:20
CYCLE COMPLETE 6:23

Digital Signature #
6AB83CD36ED333C4
