

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2029

Name: RITTER2  
Sterilizer ID: V2313228  
02/09/2023 09:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.4 F 0.0  
2:00 184.7 F 1.2  
4:00 186.7 F 5.4  
6:00 208.6 F 10.7  
8:00 226.5 F 6.1  
10:00 237.8 F 12.1  
12:00 236.9 F 10.1  
14:00 237.7 F 8.6  
16:00 250.2 F 15.9  
18:00 262.5 F 23.0  
20:00 272.9 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.5 F 30.0  
2:00 272.5 F 29.5  
3:00 271.8 F 29.0  
4:00 271.6 F 29.4  
  
Min 271.4 F 28.9  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.5 F 0.0  
5:00 220.8 F 0.0  
10:00 239.9 F 0.0  
15:00 236.4 F 0.0  
20:00 226.1 F 0.0  
25:00 222.4 F 0.0  
30:00 223.6 F 0.0  
Drying Complete

FILLING: 1:28  
HEATING: 20:20  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:57  
CYCLE COMPLETE

02/09/2023 10:50 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_