

-----  
Midmark M11 - v1.0.5  
Total Cycles: 342

Name: RITTER2  
Sterilizer ID: V2313228  
09/08/2021 01:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 174.7 F 0.0  
2:00 172.3 F 3.3  
4:00 184.0 F 7.7  
6:00 205.9 F 12.7  
8:00 218.1 F 8.1  
10:00 232.1 F 13.5  
12:00 235.5 F 11.7  
14:00 239.9 F 10.6  
16:00 252.3 F 17.1  
18:00 261.9 F 23.7  
20:00 271.4 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 30.1  
1:00 273.2 F 30.4  
2:00 272.3 F 30.8  
3:00 272.2 F 30.1  
4:00 272.4 F 30.7

Min 271.7 F 30.0  
Max 273.4 F 30.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.3 F 0.0  
5:00 237.7 F 0.0  
10:00 216.5 F 0.0  
15:00 212.5 F 0.0  
20:00 201.2 F 0.0  
25:00 196.2 F 0.0  
30:00 190.5 F 0.0  
Drying Complete

FILLING: 2:03  
HEATING: 20:57  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:00:13  
CYCLE COMPLETE

09/08/2021 02:33 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01