
Midmark M11 - v1.0.5
Total Cycles: 3245

Name: RITTER2
Sterilizer ID: V2313228
12/21/2023 12:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 101.4 F 0.0
2:00 102.0 F 0.1
4:00 119.9 F 1.5
6:00 149.2 F 4.2
8:00 172.8 F 7.3
10:00 192.3 F 10.8
12:00 215.9 F 5.8
14:00 216.4 F 8.2
16:00 226.8 F 12.5
18:00 232.6 F 8.7
20:00 238.9 F 12.1
22:00 243.1 F 12.2
24:00 254.6 F 18.2
26:00 265.3 F 24.6
28:00 270.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.5
1:00 272.1 F 29.4
2:00 272.6 F 29.6
3:00 272.7 F 29.8
4:00 272.2 F 30.0

Min 271.6 F 29.4
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.7 F 0.0
5:00 218.7 F 0.0
10:00 219.6 F 0.0
15:00 226.3 F 0.0
20:00 237.2 F 0.0
25:00 229.4 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 28:38
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:07:05
CYCLE COMPLETE

12/21/2023 01:56 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____