
Midmark M11 - v1.0.0
Total Cycles: 4007

Name: RITTER 1
Sterilizer ID: V2178198
11/10/2023 12:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.8 F 0.0
2:00 183.5 F 3.2
4:00 186.5 F 7.7
6:00 213.7 F 11.9
8:00 221.1 F 8.2
10:00 235.1 F 13.5
12:00 238.0 F 10.8
14:00 236.5 F 9.1
16:00 250.9 F 15.9
18:00 262.8 F 22.7
20:00 272.5 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.8 F 29.3
2:00 272.0 F 28.7
3:00 271.3 F 28.5
4:00 272.5 F 29.3

Min 271.2 F 28.4
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.4 F 0.0
5:00 243.9 F 0.0
10:00 224.8 F 0.0
15:00 223.1 F 0.0
20:00 223.7 F 0.0
25:00 228.4 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 20:17
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 00:59:31
CYCLE COMPLETE

11/10/2023 01:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____