

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2865

Name: RITTER 1  
Sterilizer ID: V2178198  
02/06/2023 06:44 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 207.3 F 0.0  
2:00 200.4 F 4.4  
4:00 215.0 F 10.4  
6:00 225.2 F 5.8  
8:00 239.5 F 11.8  
10:00 238.5 F 9.6  
12:00 237.0 F 8.4  
14:00 252.4 F 16.1  
16:00 265.5 F 23.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 27.8  
1:00 272.1 F 28.0  
2:00 272.7 F 28.6  
3:00 271.5 F 27.8  
4:00 272.6 F 28.4  
  
Min 271.5 F 27.7  
Max 272.8 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.3 F 0.0  
5:00 220.9 F 0.0  
10:00 219.7 F 0.0  
15:00 223.6 F 0.0  
20:00 220.9 F 0.0  
25:00 220.1 F 0.0  
30:00 217.3 F 0.0  
  
Drying Complete

FILLING: 2:55  
HEATING: 17:34  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:57:35  
CYCLE COMPLETE

02/06/2023 07:41 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_