

-----  
Midmark M11 - v1.0.5  
Total Cycles: 545

Name: RITTER2  
Sterilizer ID: V2313228  
11/11/2021 02:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 159.0 F 0.0  
2:00 158.2 F 3.2  
4:00 182.8 F 7.9  
6:00 208.6 F 12.1  
8:00 225.1 F 8.7  
10:00 236.9 F 10.0  
12:00 241.7 F 12.7  
14:00 242.0 F 11.7  
16:00 255.1 F 18.8  
18:00 267.2 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.2  
1:00 272.7 F 30.0  
2:00 272.1 F 29.3  
3:00 271.6 F 29.6  
4:00 272.7 F 30.1  
  
Min 271.2 F 29.1  
Max 272.9 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.4 F 0.0  
5:00 246.8 F 0.0  
10:00 233.0 F 0.0  
15:00 228.1 F 0.0  
20:00 213.7 F 0.0  
25:00 204.8 F 0.0  
30:00 204.6 F 0.0  
  
Drying Complete

FILLING: 1:32  
HEATING: 19:25  
STERILIZING: 4:00  
VENTING: 2:60  
DRYING: 30:00  
TOTAL CYCLE: 00:57:56  
CYCLE COMPLETE

11/11/2021 03:08 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02