

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3221

Name: RITTER 1  
Sterilizer ID: V2178198  
05/11/2023 01:22 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.2 F 0.0  
2:00 194.7 F 2.6  
4:00 210.4 F 7.9  
6:00 229.0 F 13.5  
8:00 232.9 F 9.3  
10:00 240.2 F 9.3  
12:00 245.3 F 13.1  
14:00 244.1 F 12.1  
16:00 256.9 F 18.9  
18:00 268.2 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.4 F 28.7  
2:00 271.9 F 28.6  
3:00 272.2 F 28.6  
4:00 271.6 F 28.5  
  
Min 271.4 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.7 F 0.0  
5:00 222.3 F 0.0  
10:00 235.9 F 0.0  
15:00 235.4 F 0.0  
20:00 230.7 F 0.0  
25:00 222.9 F 0.0  
30:00 221.1 F 0.0  
Drying Complete

FILLING: 2:02  
HEATING: 19:14  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:33  
CYCLE COMPLETE

05/11/2023 02:20 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_