
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002539
9:52 26/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:06
PRESSURIZING 2:11
STERILIZING 3:15
136.0 C 221kPa 3:15
136.3 C 225kPa 3:45
136.4 C 224kPa 4:15
136.3 C 225kPa 4:45
136.4 C 225kPa 5:15
136.4 C 225kPa 5:45
136.4 C 225kPa 6:15
136.5 C 225kPa 6:45
136.5 C 225kPa 6:45
VENTING 6:45
AIR DRYING 7:34
CYCLE COMPLETE 32:16

Digital Signature #
A4AD55605179D7B5
