
Midmark M11 - v1.0.5
Total Cycles: 1799

Name: RITTER2
Sterilizer ID: V2313228
12/05/2022 09:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.2 F 0.0
2:00 87.2 F 0.0
4:00 84.2 F 0.0
6:00 105.4 F 0.0
8:00 146.1 F 0.4
10:00 174.3 F 3.6
12:00 196.0 F 7.3
14:00 209.3 F 11.5
16:00 220.5 F 5.6
18:00 231.4 F 10.2
20:00 233.3 F 8.5
22:00 242.7 F 12.7
24:00 241.4 F 11.4
26:00 253.8 F 17.7
28:00 264.7 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.0 F 29.1
2:00 272.5 F 29.5
3:00 271.9 F 29.5
4:00 271.6 F 28.9

Min 271.4 F 28.8
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.5 F 0.0
5:00 230.3 F 0.0
10:00 226.0 F 0.0
15:00 224.8 F 0.0
20:00 214.9 F 0.0
25:00 211.4 F 0.0
30:00 212.4 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 29:54
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:07:46
CYCLE COMPLETE

12/05/2022 10:27 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____